

APPENDIX A

Calibration and Certification Data



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
REGION 6
MULTIMEDIA PLANNING and PERMITTING DIVISION
AIR QUALITY ANALYSIS SECTION, 6PD-Q(Houston)
10625 FALLSTONE RD.
HOUSTON, TEXAS 77099

REPORT OF ANALYSIS

Verification

The accuracy of the EPA Systems Environics 6103 photometer (Ser. # 4880) was compared to the EPA Standard Reference Photometer (SRP # 5)* on Sept. 4, 2013 by John Lay, Region 6 SRP Operator. The photometer was checked by using a protocol developed by EPA and NIST as outlined in the publication *Standard Reference Photometer for Verification and Certification of Ozone Standards*, AREAL, USEPA, RTP, NC (March, 1989).

The linear regression relationship obtained from the comparison with the EPA Region 6 SRP # 5 is given below:

$$\text{Environics 6103} = 1.0000x \text{ SRP \# 5} + 0.62 \text{ ppb}$$

A review of the comparison data indicates the tested photometer passes the verification criteria as outlined in the US EPA guidance document referenced above. Acceptable range for the slope is 0.970 to 1.030 for primary standards and 0.950 to 1.050 for transfer standards. Acceptable range for the intercept is +/- 1-3 ppb. **Note: this data is not to be used as a correction factor for primary standard data.** It is only used as a pass or fail verification of the primary standard.

*SRP # 5 is classified as a NIST Standard Reference Material (SRM) which NIST manufactured and certified.

A handwritten signature in cursive script that reads "John Lay".

John Lay
Region 6 SRP Operator,

Sept. 4, 2013

Standard Reference Photometer Calibration Report

Calibrating Institute: US EPA Region 6
Operator: John Lay
Instrument: SRP # 5 Cell Length=89.65
Comment: O3 gen by SRP
 Readings from front pannel

Date: 04-Sep-13
Start Time: 13:53
End Time: 15:14
Filename: c0904003.xls


| Calibrated Instrument: | 6103 | Calibration Results | Value | Standard Uncertainty |
|------------------------|---------------|---------------------|---------|----------------------|
| Owner: | EPA Systems | Slope | 1.00009 | 0.00133 |
| Contact: | Eric Anderson | Intercept | 0.61648 | 0.31013 |
| Make: | Envionics | Covariance | | -4.0323E-07 |
| Model: | 6103 | Res Std Dev | 0.55892 | |
| Serial Number: | 4880 | | | |

Calibration Parameters: Zero Start&End;Randomized;Raw Saved;Dark Count On (4)
Air Flow Rate: 6.5 l/min
Lamp Intensity Range: 0.0 to 58.0 %
Number Conc. Points: 8 **Points/Concentration:** 8
Conditioning: none


| Calibration Data Points | SRP # 5 | | 6103 | | 6103 | |
|-------------------------|---------|----------|--------|----------|-----------|----------|
| | Result | Std. Dev | Result | Std. Dev | Predicted | Residual |
| Dark Count 1 | 9 | | | | | |
| Dark Count 2 | 6 | | | | | |
| 1 | 0.0 | 0.4 | 1.0 | 0.8 | 0.59 | 0.41 |
| 2 | 435.4 | 0.2 | 436.0 | 1.3 | 436.08 | -0.08 |
| 3 | 16.5 | 0.1 | 17.0 | 0.1 | 17.13 | -0.13 |
| 4 | 355.1 | 0.3 | 356.0 | 0.1 | 355.76 | 0.24 |
| 5 | 272.7 | 0.3 | 274.0 | 0.1 | 273.37 | 0.63 |
| 6 | 82.1 | 0.3 | 82.0 | 0.2 | 82.72 | -0.72 |
| 7 | 187.2 | 0.1 | 187.0 | 0.2 | 187.87 | -0.87 |
| 8 | -0.1 | 0.2 | 1.0 | 0.1 | 0.48 | 0.52 |

Date: 4-Sep-13
Filename: c0904003.xls


Ozone Calibration



Residual Plot



Residual Plot



EnviroNics Calibrator Audit Flow Rate Check (As Found)



Calibrator Model : EnviroNics 6103
Serial No. : 4880
Pressure (mm Hg) 762
Temperature (F) 75
Temperature (K) 296.89
STP Correction Factor 1.00638

Date : August 26, 2013
Conducted By : EPA Systems
Flow Rate Calibrator : Bios 520 L and Bios 520 H
Serial # : 128971, 128658
Cert. Exp. Date : 7/30/2013, 9/19/2013

| Source Gas Coefficients | |
|-------------------------|------------|
| m | 1.0367E+00 |
| b | 8.7582E-02 |

| Diluent Air Coefficients | |
|--------------------------|------------|
| m | 1.0171E+00 |
| b | 9.5224E-03 |

| Source Gas | | | Diluent Air | | |
|--------------------|-----------------|-----------------|-------------------|-----------------|----------------|
| Set Point (cc/min) | Display Reading | (cc/min Actual) | Set Point (L/min) | Display Reading | (L/min Actual) |
| 5.00 | 4.94 | 5.03 | 1.000 | 0.984 | 0.986 |
| 8.00 | 7.92 | 8.20 | 1.500 | 1.489 | 1.499 |
| 10.00 | 10.0 | 10.3 | 2.000 | 1.991 | 2.007 |
| 12.00 | 12.0 | 12.4 | 2.500 | 2.491 | 2.521 |
| 15.00 | 15.0 | 15.4 | 3.000 | 2.990 | 3.031 |
| 20.00 | 20.0 | 20.7 | 3.500 | 3.490 | 3.543 |
| 25.00 | 25.0 | 25.9 | 4.000 | 3.994 | 4.054 |
| 30.00 | 30.0 | 30.9 | 4.500 | 4.493 | 4.562 |
| 35.00 | 35.0 | 36.5 | 5.000 | 4.994 | 5.075 |
| 40.00 | 39.9 | 41.6 | 5.500 | 5.493 | 5.586 |
| 45.00 | 45.0 | 47.1 | 6.000 | 5.994 | 6.095 |
| 50.00 | 49.6 | 52.1 | 6.500 | 6.493 | 6.600 |
| 60.00 | 59.9 | 62.4 | 7.000 | 6.994 | 7.118 |
| 70.00 | 69.9 | 73.0 | 7.500 | 7.493 | 7.620 |
| 80.00 | 79.9 | 83.0 | 8.000 | 7.992 | 8.120 |
| 90.00 | 89.9 | 93.0 | 8.500 | 8.493 | 8.614 |
| 100.00 | 99.9 | 103.2 | 9.000 | 8.992 | 9.143 |
| | | | 9.500 | 9.492 | 9.631 |
| | | | 10.000 | 9.991 | 10.139 |



EnviroNics Calibrator Source Gas Recalibration Flow Rate Check


Calibrator Model : EnviroNics 6103
Serial No. : 4880
Pressure (mm Hg) 762
Temperature (F) 75
Temperature (K) 296.89
STP Correction Factor 1.00638

Date : August 27, 2013
Conducted By : EPA Systems
Flow Rate Calibrator : Bios 520 L and Bios 520 H
Serial # : 128971, 128658
Cert. Exp. Date : 7/30/2013, 9/19/2013


| Source Gas Coefficients | |
|---------------------------|------------|
| <i>m</i> | <i>b</i> |
| 1.0051E+00 | 3.8511E-02 |
| Dilution Air Coefficients | |
| <i>m</i> | <i>b</i> |

| Source Gas | | | Diluent Air | | |
|--------------------|-----------------|-----------------|-------------------|-----------------|----------------|
| Set Point (cc/min) | Display Reading | (cc/min Actual) | Set Point (L/min) | Display Reading | (L/min Actual) |
| 5.00 | 4.94 | 5.00 | | | |
| 8.00 | 7.92 | 7.95 | | | |
| 10.0 | 10.0 | 10.0 | | | |
| 12.0 | 12.0 | 12.0 | | | |
| 15.0 | 15.0 | 15.0 | | | |
| 20.0 | 20.0 | 20.0 | | | |
| 25.0 | 25.0 | 25.0 | | | |
| 30.0 | 30.0 | 30.1 | | | |
| 35.0 | 35.0 | 35.4 | | | |
| 40.0 | 39.9 | 40.2 | | | |
| 45.0 | 45.0 | 45.5 | | | |
| 50.0 | 49.6 | 50.2 | | | |
| 60.0 | 59.9 | 60.2 | | | |
| 70.0 | 69.9 | 70.2 | | | |
| 80.0 | 79.9 | 80.0 | | | |
| 90.0 | 89.9 | 90.4 | | | |
| 100.0 | 99.9 | 100.7 | | | |

Source Gas



Diluent Air



CERTIFICATE OF ANALYSIS

Grade of Product: EPA Protocol

| | | | |
|------------------|---------------------|---------------------|----------------|
| Part Number: | E04NI99E15A0G9C | Reference Number: | 82-124388384-1 |
| Cylinder Number: | CC436563 | Cylinder Volume: | 144.4 CF |
| Laboratory: | ASG - Riverton - NJ | Cylinder Pressure: | 2015 PSIG |
| PGVP Number: | B52013 | Valve Outlet: | 660 |
| Gas Code: | CO,NO,NOX,SO2,BALN | Certification Date: | Sep 03, 2013 |

Expiration Date: Sep 03, 2016

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

ANALYTICAL RESULTS

| Component | Requested Concentration | Actual Concentration | Protocol Method | Total Relative Uncertainty | Assay Dates |
|-----------------|-------------------------|----------------------|-----------------|----------------------------|------------------------|
| NOX | 30.00 PPM | 31.50 PPM | G1 | +/- 1.4% NIST Traceable | 08/26/2013, 09/03/2013 |
| NITRIC OXIDE | 30.00 PPM | 31.44 PPM | G1 | +/- 1.4% NIST Traceable | 08/26/2013, 09/03/2013 |
| SULFUR DIOXIDE | 30.00 PPM | 29.85 PPM | G1 | +/- 1.1% NIST Traceable | 08/26/2013, 09/03/2013 |
| CARBON MONOXIDE | 3000 PPM | 3001 PPM | G1 | +/- 1.0% NIST Traceable | 08/29/2013 |
| NITROGEN | Balance | | | | |

CALIBRATION STANDARDS

| Type | Lot ID | Cylinder No | Concentration | Uncertainty | Expiration Date |
|----------|--------------|--------------|--|-------------|-----------------|
| NTRMplus | 12061636 | CC344928 | 20.23 PPM NITRIC OXIDE/NITROGEN | +/- 0.9% | Apr 11, 2015 |
| PRM | 12312 | 680179 | 10.01 PPM NITROGEN DIOXIDE/NITROGEN | +/- 2.0% | Feb 14, 2012 |
| NTRMplus | 12060809 | CC281027 | 49.95 PPM NITRIC OXIDE/NITROGEN | +/- 0.8% | Dec 16, 2017 |
| GMIS | 124206889106 | CC322664 | 4.879 PPM NITROGEN DIOXIDE/NITROGEN | +/- 2.0% | Apr 08, 2016 |
| NTRM | 12061832 | CC352180 | 50.10 PPM SULFUR DIOXIDE/NITROGEN | +/- 1.0% | Apr 24, 2018 |
| NTRM | 000525 | SG9152275BAL | 1985 PPM CARBON MONOXIDE/NITROGEN | +/- 0.6% | Aug 17, 2016 |

The SRM or PRM noted above is only in reference to the GMIS used in the assay and not part of the analysis.

ANALYTICAL EQUIPMENT

| Instrument/Make/Model | Analytical Principle | Last Multipoint Calibration |
|-----------------------------------|----------------------|-----------------------------|
| Siemens Ultramat 6 N1C8180 COHIGH | NDIR | Aug 09, 2013 |
| Nicolet 6700 APW1100391 NO | FTIR | Aug 23, 2013 |
| Nicolet 6700 APW1100391 NO2 | FTIR | Aug 23, 2013 |
| Nicolet 6700 APW1100391 SO2 | FTIR | Aug 10, 2013 |

Triad Data Available Upon

Notes:



Approved for Release

Airgas Specialty Gases

600 Union Landing Road

Cinnaminson, NJ 08077

(856) 829-7878 Fax: (856) 829-6576

www.airgas.com